

Notice of Allowability

Application No.

09/824,074

Examiner

Charles G Freay

Applicant(s)

DOVEY ET AL.

Art Unit

3746

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to the Request for Continuing Examination and the Preliminary Amendment of April 12, 2004.
2. ☒ The allowed claim(s) is/are 1-8.
3. ☒ The drawings filed on 15 November 2002 are accepted by the Examiner.
4. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☒ All b) ☐ Some* c) ☐ None of the:
 1. ☒ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.


Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 6. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☒ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☒ Information Disclosure Statements (PTO-1449 or PTO/SB/08),
Paper No./Mail Date 04122004
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application (PTO-152)
6. ☐ Interview Summary (PTO-413),
Paper No./Mail Date _____.
7. ☐ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____.


Charles G Freay
Primary Examiner
Art Unit: 3746

REASONS FOR ALLOWANCE

The following is an examiner's statement of reasons for allowance: the prior art neither discloses nor makes obvious an apparatus having a reciprocating piston, a drive means and a cylinder as claimed including a sensor means in communication with the cylinder for sensing contact of the piston and a control means interconnecting the sensor means and the drive means, the control means is adapted to receive the contact signal as a sole input signal and generate a control signal to the drive means to adjust reciprocating movement of the piston.

Prior art references, such as the previously cited Yang reference and the newly cited Unger et al reference, which disclose a linear reciprocating piston device with a collision sensor disclose the motor controller having multiple inputs and feed back parameters. Removing the additional inputs and controls would detrimentally effect the operation of the devices and there is no reason to do so. Other prior art devices, such as Matsumura, Lilie et al and Schwarz et al disclose apparatus with piston position sensors. In these devices the sensing and control is specifically designed to avoid any piston/cylinder contact. Therefore it would not be obvious to modify these devices with collision sensor devices and controls.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Art Unit: 3746

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Charles G Freay whose telephone number is 703-308-0639. The examiner can normally be reached on Monday through Friday 10:00 A.M. to 5:30 P.M..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Justine Yu can be reached on 703-308-2675. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Charles G Freay
Primary Examiner
Art Unit 3746

CGF
July 7, 2004